



HYALURE LIGHT

Filler for tissue augmentation based on hyaluronic acid and Lidocaine

Manufacturer: **Bioitech S.r.l.** - Via Chiana, 87 - 00198 Roma

INSTRUCTIONS FOR USE

Composition

Hyalure Lips: Crosslinked Hyaluronic Acid 20 mg/g, Lidocaine 3 mg/g, Saline Phosphate Buffer pH 7 q.s. 1 mL.

Description and presentation

Hyalure is a sterile intradermal injection system, based on Sodium hyaluronate, of biotechnological origin, thus of non-human, nonanimal origin.

Hyalure adopts a crosslinking process to improve its stay in the human body.

Hyalure is sold in disposable pre-filled syringes. Each pack contains 2 syringes, 4 disposable sterile needles for the injection of the product, a package insert and 2 traceability labels containing the lot number of the device.

Indications

Hyalure is a product intended exclusively for medical personnel trained in the infiltration of biomaterials for dermatological, reconstructive and aesthetic use (filling).

Hyalure is fully biodegradable and slowly resorbable. A few sessions at regular intervals helps to maintain the result of the correction.

Hyalure Lips with 30 G ½" needle is suitable for: filling vermilion and lips.

Contraindications

Hyalure cannot be used:

- for the treatment of periocular wrinkles
- for intravascular injection
- in association with peeling, laser, dermabrasion
- in patients with a proven hypersensitivity to product components
- in pregnant and nursing women
- on persons under the age of 18
- in patients with a proven tendency to develop hypertrophic scars

- on burned skin or the skin of patients undergoing radiotherapy treated with acetylsalicylic acid and anticoagulants
- in the presence of inflammation or infection of the area to be treated.

Side effects

The doctor must inform the patient about any undesirable side effects related to the implanting of Hyalure. Side effects may occur immediately after injection or after a certain period of time.

Temporary erythemas in the infiltrated area, which resolve within 36 hours, are possible with use of the product.

Other side effects may include:

- inflammatory reactions and edemas, which may be associated with itching
- pain immediately after implanting
- hardening or nodules in the injection site
- coloration of the skin in the injection site
- low effectiveness of the treatment
- insufficient filling effect.

Furthermore, resorbable fillers have been associated on rare occasions with secondary phenomena such as: necrosis of the glabellar region, abscesses, fistulas and granulomas of the treated area.

The persistence of inflammatory reactions for more than 7 days or the occurrence of side effects other than those described should be promptly communicated to the physician, who will immediately report them to the distributor or manufacturer.

Precautions

Hyalure is only suitable for intradermal and subcutaneous injections.

Do not inject into an area that has already been injected with a non resorbable filler or that has recently been treated.

The use of cosmetics containing alcoholic solutions is not recommended in the hours following the infiltration of Hyalure.

Exposure to sunlight, UV rays, intense cold and the use of saunas or steam baths should be avoided for 48h.

Prevent the contents of the syringe from coming into contact with solutions based on Quaternary Ammonium.

Any residual air between the cap and the syringe body is normal and should not be considered as a sign of alteration of the product.

Throw away the device and the remains of the product after use.

Dosage and administration

The doctor's mastering of injection techniques is essential for successful treatment. The patient must be informed about possible adverse effects that might arise and contraindications of the treatment.

The injection site must be thoroughly disinfected before treatment. The product should be injected slowly with the needle supplied. The amount of Hyalure to be injected and the injection mode depend on the type of wrinkle being treated.

It is advisable to massage the treated area after injection to spread the product more evenly within the tissues.

Overcorrection of the areas to be treated is not recommended.

Use of Hyalure device:

- Open the package containing a pre-filled syringe
- Remove the cap on the luer-lock connector of the syringe by unscrewing
- Open the package containing a needle
- Screw the needle clockwise on the luer-lock connector of the syringe without removing the cap
- Remove the needle cap
- Push the syringe plunger to remove air until the increase in resistance to pressure is felt
- Apply the product by multiple injections after having thoroughly disinfected the area to be treated
- The syringe scale is indicative, and serves to better dose the product during application.

Warnings and storage conditions

- **This product is for the exclusive use of trained medical personnel**
- Follow the instructions for use
- Check the integrity of the package and the syringe before use
- Check the expiration date on the package
- Do not use the product after the expiration date
- Do not resterilize
- Use the contents of the syringe and the needle on a single patient
- Do not reuse
- Used needles should be discarded in special containers

- Keep away from high and low temperatures
- Avoid sun exposure
- Avoid knocks
- Store between 8°C and 25°C.

Package: 2 disposable pre-filled syringes, 4 disposable sterile needles.